

Premi® Test ampoule photo instructions



1. Cut off the required number of ampoules with a pair of scissors; be careful not to damage the foil of remaining ampoules.



5. Flush the meat juice away by washing the test twice with demi water and carefully remove the water from the test.



2. Take approximately 2 cm³ of lean meat and use a meat press to extract about 250 µl of meat juice.



6. Close the test ampoule with foil.



3. Pipette 100 µl of meat juice slowly onto the agar in the ampoule. Do not distort the agar.



7. Preheat the DSM heating block incubator at the required temperature of 64.0°C (or 147°F). Put the ampoule in the Incubator for approximately 3 hours.



4. Allow to stand at room temperature for 20 minutes for a prediffusion.



8. Withdraw the ampoule from the block heater after the negative control changed color.



9. Put the ampoules in the template on the scanner.



10. Read the color of the ampoules with the Premi® Scan.

Version: 0507
Date of issue: March 17, 2006

DSM Nutritional Products
DSM Premi® Test B.V., P.O. Box 1163,
6160 BD Geleen, The Netherlands, Tel. +31 (0)46 4763573
Fax. +31 (0)46 4763737, Internet site: www.premitest.com

Although the facts and suggestions in this publication are based on our own research and are believed reliable, we cannot assume any responsibility for performance or results obtained through the use of our products herein described, nor do we accept any liability for loss or damages directly or indirectly caused by our products. The user is held to check the quality, safety and all other properties of our product prior to use. Nothing herein is to be taken as permission, inducement or recommendation to practise any patented invention without a license.